

Product approval

CERTIFICATE NO.: NET527P04

Holder of certificate: KiiltoClean AS, Norway

Producer: KiiltoClean AS
Hegaløkkveien 13,
N-1383 Asker
NORWAY

Product: Antibac® Manuelle Ståldispensere [manual steel dispenser].

Identification: Article no. 601003 tested. Steel dispensers with the same combinations of materials also approved.

Dimensions: 173 x 112 x 88 mm

Test laboratory: Nordisk Emballasje Testing AS, Norway

Test standard: ITP-214, Desinfeksjon. V1.

The product approval is based on an evaluation of three samples of the above-described product. The test results relate only to the items tested.

Test report: NET527T04

Appendix: Page 2 in this document

Brevik, 24.11.2015



**Nordisk Emballasje Testing AS
Certification**

Certified by:



Rune Madsen Fink
Managing Director

APPENDIX TO CERTIFICATE NET527P04

Considering usage in circumstances where hygiene and disinfection are both important factors, this product has been evaluated and tested by NET. The Antibac® Manuelle Ståldispensere has been exposed to sterilizing temperatures and chemicals.

The following tests have been performed:

Test description	Temperature	Result
Autoclaving, 15 minutes	121°C	OK, no visual damage
Cleaning and disinfection with Antibac® Overflatedesinfeksjon 75%	23°C	OK, no visual damage
Cleaning with normal detergent	23°C	OK, no visual damage
Decontamination with steam, 15 minutes	85°C	OK, no visual damage

NOTE: The dispenser series have a body of galvanized steel, and rods for both holding and dispensing disinfectant or soap in a magnetic alloy. The galvanized body is not magnetic.

