

## Product approval

**CERTIFICATE NO.: NET527P05**

**Holder of certificate:** KiiltoClean AS, Norway

**Producer:** KiiltoClean AS  
Hegaløkkveien 13,  
N-1383 Asker  
NORWAY

**Product:** Antibac® Berøringsfri Dispenser Hvit [touch-less dispenser system, white].

**Identification:** Article no. 600861 tested. Automatic, battery driven dispenser for bags of disinfectants, soap etc.

**Dimensions:** 265 x 110 x 115 mm

**Test laboratory:** Nordisk Emballasje Testing AS, Norway

**Test standard:** ITP-214, Dekontaminering. V1

The product approval is based on an evaluation of three samples of the above-described product. The test results relate only to the items tested.

**Test report:** NET527T05

**Appendix:** Page 2 in this document

**Brevik, 24.11.2015**



**Nordisk Emballasje Testing AS**  
**Certification**

**Certified by:**



**Rune Madsen Fink**  
Managing Director

## APPENDIX TO CERTIFICATE NET527P05

Considering usage in circumstances where hygiene and disinfection are both important factors, this product has been evaluated and tested by NET. The Antibac® Berøringsfri Dispenser Hvit has been exposed to sterilizing temperatures and chemicals.

The following tests have been performed:

| Test description  | Temperature | Result               |
|---|-------------|----------------------|
| Cleaning with normal detergent                                    | 23°C        | OK, no visual damage |
| Cleaning and disinfection with Antibac® Overflatedesinfeksjon 75% | 23°C        | OK, no visual damage |
| Decontamination with steam, 15 minutes                            | 85°C        | OK, no visual damage |

**NOTE:** These automatic dispensers have a body of plastics and a battery pack, which was removed for the purpose of steam-testing.

