

Product approval

CERTIFICATE NO.: NET527P06

Holder of certificate:	KiiltoClean AS, Norway
Producer:	KiiltoClean AS Hegaløkkveien 13, N-1383 Asker NORWAY
Product:	Antibac™ Softbagdispenser.
Identification:	Article no. 601001 tested. Plastics, wall mounted dispenser for Softbags containing disinfectants, soap etc.
Dimensions:	220 x 195 x 100 mm
Test laboratory:	Nordisk Emballasje Testing AS, Norway
Test standard:	ITP-214, Dekontaminering. V1

The product approval is based on an evaluation of three samples of the above-described product. The test results relate only to the items tested.

Test report:	NET527T06
Appendix:	Page 2 in this document

Brevik, 24.11.2015



**Nordisk Emballasje Testing AS
Certification**

Certified by:


Rune Madsen Fink
Managing Director

APPENDIX TO CERTIFICATE NET527P06

Considering usage in circumstances where hygiene and disinfection are both important factors, this product has been evaluated and tested by NET. The Antibac™ Softbagdispenser has been exposed to sterilizing temperatures and chemicals.

The following tests have been performed:

Test description	Temperature	Tested item	Result
Autoclaving, 15 minutes	121 °C	Handle & front panel	OK, no visual damage
Cleaning with normal detergent	23 °C	Complete dispenser	OK, no visual damage
Cleaning and disinfection with Antibac™ Overflatedesinfeksjon 75%	23 °C	Complete dispenser	OK, no visual damage
Decontamination with steam, 15 minutes	85 °C	Complete dispenser	OK, no visual damage

